

Package ‘OSLdecomposition’

October 12, 2022

Type Package

Title Signal Component Analysis for Optically Stimulated Luminescence

Version 1.0.0

Date 2022-07-29

Maintainer Dirk Mittelstraß <dirk.mittelstrass@luminescence.de>

Description Function library for the identification and separation of exponentially decaying signal components in continuous-wave optically stimulated luminescence measurements.

A special emphasis is laid on luminescence dating with quartz, which is known for systematic errors due to signal components with unequal physical behaviour.

Also, this package enables an easy to use signal decomposition of data sets imported and analysed with the R package 'Luminescence'.

This includes the optional automatic creation of HTML reports. Further information and tutorials can be found at <<https://luminescence.de>>.

Depends R (>= 4.0)

License GPL-3

BugReports <https://github.com/DirkMittelstrass/OSLdecomposition/issues>

Encoding UTF-8

RoxygenNote 7.2.0

Imports Luminescence (>= 0.9.9), methods, utils, stats, tools,
DEoptim, minpack.lm, gridExtra, ggplot2, scales, ggpubr,
rmarkdown

Suggests kableExtra

NeedsCompilation no

Author Dirk Mittelstraß [aut, cre] (<<https://orcid.org/0000-0002-9567-8791>>),
Sebastian Kreuzer [aut] (<<https://orcid.org/0000-0002-0734-2199>>),
Christoph Schmidt [aut] (<<https://orcid.org/0000-0002-2309-3209>>),
Jan Beyer [ths] (<<https://orcid.org/0000-0002-1403-395X>>),
Johannes Heitmann [ths],
Arno Straessner [ths] (<<https://orcid.org/0000-0003-2460-6659>>)

Repository CRAN

Date/Publication 2022-08-12 12:30:02 UTC

R topics documented:

OSLdecomposition-package	2
decompose_OSLcurve	3
fit_OSLcurve	7
optimise_OSLintervals	11
plot_OSLcurve	13
plot_PhotoCrosssections	16
RLum.OSL_correction	18
RLum.OSL_decomposition	21
RLum.OSL_global_fitting	24
simulate_OSLcomponents	28
sum_OSLcurves	30
Index	32

OSLdecomposition-package

Signal Component Analysis for Optically Stimulated Luminescence

Description

Function library for the identification and separation of exponentially decaying signal components in continuous-wave optically stimulated luminescence (CW-OSL) measurements. A special emphasis is laid on luminescence dating with quartz, which is known for systematic errors due to CW-OSL signal components with unequal physical behaviour. Also, this package enables an easy to use signal decomposition of CW-OSL data sets imported and analysed with the R package 'Luminescence'. This includes the optional automatic creation of HTML reports.

Details

Project website

- <https://luminescence.de>

Source code repository

- <https://github.com/DirkMittelstrass/OSLdecomposition>

Bug reporting

- <https://github.com/DirkMittelstrass/OSLdecomposition/issues>

This package is part of the RLum.Network

- <https://r-luminescence.org>

Package maintainer

Dirk Mittelstraß, Independent researcher, Dresden (Germany),
<dirk.mittelstrass@luminescence.de>

Citation

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

Funding Dirk Mittelstraß created this package as part of his master thesis and further enhanced and published it as private endeavour. He did not receive any specific grant from funding agencies in the public, commercial or not-for-profit sectors.

decompose_OSLcurve *Multi-exponential CW-OSL decomposition*

Description

Function for determining the signal component amplitudes of a multi-exponential decay curve if the signal component decay parameters are already given. Thus, this function decomposes CW-OSL curves with known components of unknown intensity.

Usage

```
decompose_OSLcurve(
  curve,
  components,
  background.fitting = FALSE,
  algorithm = "det",
  error.estimation = "empiric",
  verbose = TRUE
)
```

Arguments

curve	data.frame or matrix or RLum.Data.Curve (required): CW-OSL curve x-Axis: \$time or first column as measurement time (must have constant time intervals); y-Axis: \$signal or second column as luminescence signal. Further columns will be ignored.
components	data.frame or numeric vector (required): Either a vector containing the decay parameters of the CW-OSL components or a table (data.frame), usually the table returned by fit_OSLcurve . In case of a vector: It is recommended to use less than 7 parameters. The parameters will be sorted in decreasing order. In case of a data.frame, one column must be named \$lambda. It is recommended to provide also integration interval parameters (columns \$t.start, \$t.end, \$ch.start, \$ch.end), which can be found by applying optimise_OSLintervals to the global mean curve, calculated by sum_OSLcurves . If one or more column is missing, a simple interval definition algorithm is run automatically, see section Details .

background.fitting	logical (<i>with default</i>): if TRUE, an additional signal component with a decay rate of $\lambda = 0$ is included. This allows for an accurate estimation of slow component intensities if the data is not background corrected. However, the additional component reduces the overall precision of the algorithm. It can also cause implausible slow component results if the measurement duration is not sufficiently long (> 30 s).
algorithm	character string (<i>with default</i>): Choice of curve decomposition approach. Either "det" or "det+nls" or "nls", see section Details . ^^
error.estimation	character string (<i>with default</i>): integral error estimation approach, either "empiric" or "poisson" or a numeric value or "none", see section Details . This argument has no effect if algorithm = "nls".
verbose	logical (<i>with default</i>): enables console text output

Details

The function assumes multiple exponentially decaying signal components with first-order kinetics:

$$I(t) = n_1 \lambda_1 \exp(-\lambda_1 t) + n_2 \lambda_2 \exp(-\lambda_2 t) + \dots + n_K \lambda_K \exp(-\lambda_K t)$$

with $I(t)$ the CW-OSL signal, n the signal component intensity, λ the signal component decay constant and K the number of signal components. For the actual decomposition procedure, the integrated version of this formula is used, see Mittelstrass et al. (2021) for details.

Decomposition algorithm

The calculation procedure depends on the function argument `algorithm`. This function includes two different decomposition algorithms: "det" for **d**eterminant solution and "nls" for **n**onlinear least squares estimate

`algorithm = "det"` (default)

The function calculates the CW-OSL component intensities by building an equation system which is then solved by a determinant-based approach (Cramers rule). This purely analytical approach gives the algorithm a solution in all possible cases, even if the measurement consists just of noise or the wrong model is used. There are also no 'false minima' events. The statistical error is calculated by applying the *propagation of uncertainty* method on Cramers rule.

The precision of this algorithm as well as the propagation of eventual systematic errors of the decay rate values, depend on the integration intervals, given by the columns `$t.start`, `$t.end`, `$ch.start` and `$ch.end` of the `data.frame` used as input for the argument components. In principle, these can be chosen freely. Reasonable integration intervals are defined by `optimise_OSLintervals`. If not defined, the logarithmic mean values between life times (reciprocal decay rate) of subsequent components are used as interval borders.

`algorithm = "nls"`

As alternative algorithm, Levenberg-Marquardt nonlinear regression is available, see `minpack.lm::nlsLM` for details. The results are identical to that of the "det" algorithm in accuracy and precision. But there is the slight chance (< 1 %) of fitting failure when using the "nls" algorithm. Also, the statistical errors are underestimated by 20-80 % in most cases. As advantage, the "nls" algorithm is less sensitive against systematic errors caused by uncorrected signal background.

algorithm = "det+nls"

Both algorithms can be combined. Then, "det" provides the starting values and the error estimations for "nls" and returns replacement results, in case "nls" fails. "nls" compensates for potential systematic errors in the fast and medium components intensity values due to uncorrected signal background. However, the background signal will still affect slow component results. The slowest component will be overestimated while the second slowest component will be underestimated. If these components are of particular interest, it is recommended to set `background.fitting = TRUE`

All three methods were tested at 5×10^6 simulated CW-OSL curves by Mittelstrass (2019) for their performance (+++ reliable results in all cases; ++ reliable in >95% of cases; + reliable in most cases):

Characteristics	det	nls	det+nls
Decomposition success rate	100 %	>99 %	100 %
Component intensity accuracy	+++	+++	+++
Accuracy in case of uncorrected background	+	++	++
Error estimation accuracy	+++	+	++

In summary, algorithm = "det" is recommended for the most cases. If the signal background level is significant (> 2 % of initial signal) but was not corrected, algorithm = "det+nls" is the better choice. Setting `background.fitting = TRUE` is usually not recommended, only in case slow components shall be investigated in measurements with uncorrected background.

Error estimation

In case of algorithm = "det" or "det+nls" the Propagation of Uncertainty method is used to transform signal bin error values (column `$bin.error`) into component intensity error values (column `$n.error`). The signal bin error calculation depends on the argument `error.estimation`, see below. If algorithm = "nls" is used, the error values provided by `minpack.lm::nlsLM` are returned.

`error.estimation = "empiric"` (default)

The standard deviation of each signal bin (signal bin = signal value of an integrated time interval) is calculated from the *corrected sample variance* between the CW-OSL model and the actual CW-OSL curve for that interval. Thus, statistical errors are monitored accurately without any prior knowledge required. However, potential systematic errors are monitored insufficiently. Also, at least two (better more) data points per signal bin are needed to estimate its standard deviation. If a signal bin consists just of one data point, its square root value is taken as standard deviation, in accordance to the Poisson distribution.

`error.estimation = "poisson"` or `numeric` value

Alternatively the standard error can be calculated by approximating a **Poisson** distributed signal error, known as *Shot noise*. This is suitable if the lack of data points on the x-axis circumvents an empiric error estimation, like with spatially or spectrally resolved CCD measurements. Also the parameter can be set to a `numeric` value, which represents the detector noise in *cts/s* and is assumed to be normal distributed. The detector noise will be added on top of the Poisson distributed shot noise.

`error.estimation = "only.bin.RSS"`

The error estimation is omitted but the residual sum of squares (RSS) between input curve and combined signal component curves is calculated. However, the RSS value is divided into sections

according to the signal bins (column `$bin.RSS`). The full RSS value can be calculated by summing over the complete column. The RSS value is usually used a minimization target in fitting algorithms, like done in `fit_OSLcurve`. The values of the `$bin.RSS` column allows for weighted fitting by applying pre-factors to the bin RSS values. For further speed advance, the calculation of `$components$n.residual` and `$components$initial.signal` is also omitted.

```
error.estimation = "none"
```

The error estimation is omitted. This option saves significant computing time, if the error estimation is not of significance. For further speed advance, the calculation of `$components$n.residual` and `$components$initial.signal` is also omitted.

Systematic errors

The ratio of the error values of both error estimation methods can be used to detect (but not quantify) systematic errors. "poisson" error values are not affected by systematic errors, while "empiric" errors are. If the detector noise is known and taken into account, the relation between both values for a given signal bin should be about $empiric/poisson = 1$. In case of systematic errors, this ratio increases.

Value

The input table `components data.frame` will be returned with added or overwritten columns: `$n`, `$n.error`, `$n.residual`, `$bin`, `$bin.error`, `$bin.RSS`, `$initial.signal`. Which columns are written depends on the selected parameters. If an input `data.frame` contains already one of the above columns but parameters are selected which do not re-calculate the values, the values of the columns are set to NA.

Last updates

2022-07-25, DM: Extended algorithm for bin-wise RSS calculation and added error estimation option "only.bin.RSS"

Author(s)

Dirk Mittelstraß, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Mittelstraß, D., 2019. Decomposition of weak optically stimulated luminescence signals and its application in retrospective dosimetry at quartz (Master thesis). TU Dresden, Dresden.

See Also

[fit_OSLcurve](#), [optimise_OSLintervals](#), [RLum.OSL_decomposition](#), [minpack.lm::nlsLM](#)

Examples

```
# Set some arbitrary decay parameter for a dim CW-OSL measurement of quartz
components <- data.frame(name = c("fast", "medium", "slow"),
                        lambda = c(2, 0.5, 0.02),
                        n = c(1000, 1000, 10000))

# Simulate the CW-OSL curve and add some signal noise and some detection background
curve <- simulate_OSLcomponents(components, simulate.curve = TRUE,
                               add.poisson.noise = TRUE, add.background = 40)

# Decompose the simulated curve
components <- decompose_OSLcurve(curve, components)

# Display the component separation results
plot_OSLcurve(curve, components)

### Decomposition including signal background fitting:

# Define optimized integration intervals, including an interval for the background
components <- optimise_OSLintervals(components, curve, background.component = TRUE)

# Decompose again and view results
components <- decompose_OSLcurve(curve, components, background.fitting = TRUE)
plot_OSLcurve(curve, components)
```

fit_OSLcurve

Multi-exponential CW-OSL curve fitting

Description

Fitting function for multi-exponentially decaying CW-OSL measurements, based on the algorithm described by Bluszcz & Adamiec (2006).

Usage

```
fit_OSLcurve(
  curve,
  K.max = 5,
  F.threshold = 150,
  stimulation.intensity = 30,
  stimulation.wavelength = 470,
  verbose = TRUE,
  output.complex = FALSE,
  parallel.computing = FALSE
)
```

Arguments

curve	RLum.Data.Curve or data.frame or matrix (required): CW-OSL record or average CW-OSL curve created by sum_OSLcurves . If no column <code>\$time</code> exists, the first column is defined as measurement time (x-axis). Time intervals must be constant. If no column <code>\$signal</code> exists, the second column is defined as signal values (y-axis). Further columns will be ignored
K.max	numeric (<i>with default</i>): Maximum number of components K . The computing time increases exponentially with the component number. $K < 7$ is recommended
F.threshold	numeric (<i>with default</i>): Fitting stop criterion. If the F-value is lower than this threshold, the fitting procedure stops and the $K - 1$ fit is returned
stimulation.intensity	numeric (<i>with default</i>): Intensity of optical stimulation in mW / cm^2 . Used to calculate photoionisation cross sections.
stimulation.wavelength	numeric (<i>with default</i>): Wavelength of optical stimulation in nm . Used to calculate photoionisation cross sections. If a wavelength between 465 and 480 nm is chosen, the cross sections are set into relation with literature values to name the signal components automatically.
verbose	logical (<i>with default</i>): Enables console text output.
output.complex	logical (<i>with default</i>): If TRUE, the function returns a list of objects, see section Value for further information. If FALSE, the function returns a data.frame with the CW-OSL model parameters of the fitting chosen by the F-test. Setting the parameter to FALSE is not recommended when fitting a global average curve created by sum_OSLcurves as over-fitting is likely in such cases.
parallel.computing	logical (<i>with default</i>): Enables the use of multiple CPU cores. This increases the execution speed significantly but may need administrator rights and/or a firewall exception. See DEoptim::DEoptim.control for further information.

Details

The function assumes multiple exponentially decaying signal components with first-order kinetics:

$$I(t) = n_1 \lambda_1 \exp(-\lambda_1 t) + n_2 \lambda_2 \exp(-\lambda_2 t) + \dots + n_K \lambda_K \exp(-\lambda_K t)$$

with $I(t)$ the CW-OSL signal, n the signal component intensity, λ the signal component decay constant and K the number of signal components. For actual fitting, the integrated version of this formula is used, see Mittelstraß et al. (2021) for details.

The fitting algorithm is an implementation of the *hybrid evolutionary-linear algorithm* (HELA) by Bluszcz & Adamiec (2006). See there or Mittelstraß et al. (in preparation) for details. The differential evolution part of HELA is performed by [DEoptim::DEoptim](#). The linear regression part of HELA is performed by [decompose_OSLcurve](#). The parameter refinement by Levenberg-Marquardt fitting is performed by [minpack.lm::nlsLM](#).

F-test

Bluszcz & Adamiec (2006) suggest the use of an F-test to determine the correct number of signal components. This function compares the residual square sum (RSS_K) value of each fitting with the RSS_{K-1} value of the previous fitting and calculates an *Improvement-in-fitting-quality* criterion:

$$F_K = (RSS_{K-1} - RSS_K)/2/RSS_K(N - 2K)$$

Here, N is the number data points (channels) of the measurement and K is the number of OSL components in the fitting model. If F_K falls below the threshold value (`F.threshold`), the fitting model with K components is apparently not significantly better than the $K - 1$ model of the previous fitting cycle. Thus, the $K - 1$ model will be recommended as fitting solution.

Photoionisation cross sections

While the function is suited for the analysis of a wide variety of multi-exponential decay problems, it is targeted to CW-OSL measurements of quartz under SAR protocol conditions (470 nm stimulation at 125 °C). To compare the calculated OSL components with OSL components reported in published literature, photoionisation cross sections are calculated using the stimulation.wavelength λ_{stim} and stimulation.intensity Φ_{stim} :

$$\sigma_k = \lambda_k hc / \Phi_{stim} \lambda_{stim}$$

Here σ_k is the photoionisation cross section of component k in cm^2 , λ_k the CW-OSL decay constant in s^{-1} , h the Planck constant and c the speed of light.

If a stimulation.intensity between 460 nm and 485 nm is defined, the components are named automatically in accordance to the cross-sections published by Durcan and Duller (2011), Jain et al. (2003) and Singarayer and Bailey (2003). For the Ultrafast and the Slow4 component, no consistent literature values could be found, so their range is tentatively assigned:

Component	Lower limit (cm^2)	Upper limit (cm^2)
Ultrafast	1e-16	1e-15
Fast	1.9e-17	3.1e-17
Medium	3e-18	9e-18
Slow1	1e-18	1.85e-18
Slow2	1.1e-19	4e-19
Slow3	1e-20	4.67e-20
Slow4	1e-21	1e-20

Value

If `output.complex = FALSE`, a [data.frame](#) is returned. It contains the signal decay rates and signal intensities of the best fit. The best fit was either chosen by the F-test or the last successful fitting iteration.

If `output.complex = TRUE`, a [list](#) of objects is returned:

Element	Type	Description
<code>decay.rates</code>	numeric	vector of the best suiting decay rates
<code>K.selected</code>	numeric	number of components of the best fit
<code>F.test</code>	data.frame	table containing the F-test parameter and the decay rates of each fitting model
<code>F.test.print</code>	data.frame	the same table as above, but formatted for pretty console and report output

info.text	character	collected messages from the algorithms
component.tables	list	result data.frames for all tested models
curve	data.frame	fitted time-signal-curve
components	data.frame	best fit; same object as <code>output.complex = FALSE</code> returns
fit.results	list	list of nls objects for all tested models
plot.data	data.frame	factorized results for overview plotting with plot_PhotoCrosssections
parameters	list	function arguments and the needed computing time

Last update

2022-07-27, DM: Moved residual sum of squares (RSS) calculation during DE-optimization cycle to `decompose_OSLcurve()` to improve computing time by factor 3 to 4

Author(s)

Dirk Mittelstraß, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Bluszcz, A., Adamiec, G., 2006. Application of differential evolution to fitting OSL decay curves. *Radiation Measurements* 41, 886–891.

Durcan, J.A., Duller, G.A.T., 2011. The fast ratio: A rapid measure for testing the dominance of the fast component in the initial OSL signal from quartz. *Radiation Measurements* 46, 1065–1072.

Jain, M., Murray, A.S., Bøtter-Jensen, L., 2003. Characterisation of blue-light stimulated luminescence components in different quartz samples: implications for dose measurement. *Radiation Measurements* 37, 441–449.

Mittelstraß, D., 2019. Decomposition of weak optically stimulated luminescence signals and its application in retrospective dosimetry at quartz (Master thesis). TU Dresden, Dresden.

Singarayer, J.S., Bailey, R.M., 2003. Further investigations of the quartz optically stimulated luminescence components using linear modulation. *Radiation Measurements, Proceedings of the 10th international Conference on Luminescence and Electron-Spin Resonance Dating (LED 2002)* 37, 451–458.

See Also

[RLum.OSL_decomposition](#), [sum_OSLcurves](#), [decompose_OSLcurve](#), [plot_OSLcurve](#), [plot_PhotoCrosssections](#), [minpack.lm::nlsLM](#), [DEoptim::DEoptim](#)

Examples

```
# Create a simple curve with just one component
curve <- data.frame(
  X = c(1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 11, 12),
```

```

Y = c(377, 244, 163, 93, 59, 28, 17, 13, 10, 8, 9, 5))
# Perform fitting
components <- fit_OScurve(curve, F.threshold = 3)

# Display results
plot_OScurve(curve, components)

```

optimise_OSIntervals *Find adequate integration intervals for CW-OSL decomposition*

Description

This function defines integration intervals for CW-OSL component separation with [decompose_OScurve](#). The underlying iterative optimisation process aims for minimum cross-correlation between the signal components.

Usage

```

optimise_OSIntervals(
  components,
  curve = NULL,
  channel.width = NA,
  channel.number = NA,
  t.start = 0,
  t.end = NA,
  background.component = FALSE,
  verbose = TRUE,
  parallel.computing = FALSE
)

```

Arguments

components	data.frame or numeric vector (required): Table or vector containing the decay constants of the signal components. A data.frame must contain a column λ . Usually the data.frame is provided by fit_OScurve .
curve	data.frame or matrix or RLum.Data.Curve (<i>optional</i>): OSL signal curve which serves as time axis template. The input curve will be used to define <code>channel.width</code> and <code>channel.number</code>
channel.width	numeric (<i>optional</i>): Channel width in seconds. Necessary if <code>curve</code> is not given.
channel.number	numeric (<i>optional</i>): Number of channels resp. data points. Necessary if <code>curve</code> is not given.
t.start	numeric (<i>with default</i>): Starting time of the first interval, per default the start of the measurement.
t.end	numeric (<i>optional</i>): End time of the last interval, per default the end of the measurement.

background.component

logical (*with default*): If TRUE, an additional interval for a component with a decay rate of zero will be determined. This enables the calculation of the signal background level during the signal decomposition with [decompose_OSLcurve](#).

verbose

logical (*with default*): Enables console text output.

parallel.computing

logical (*with default*): Enables the use of multiple CPU cores. This increases the execution speed significantly but may need administrator rights and/or a firewall exception. See [DEoptim::DEoptim.control](#) for further information.

Details

The precision of the component separation with [decompose_OSLcurve](#) and the impact of systematic decay rate errors on the component separation depends on the integration interval definition. This function minimises the influence of an under/over-estimated decay rate to the signal intensity calculation of other component. This is done by maximizing the denominator determinant in Cramers rule, see Mittelstraß (2019) for details. For maximisation, the iterative evolutionary algorithm of Storn and Price (1997) is used, available in R through [DEoptim::DEoptim](#).

The inclusion of a background component is supported, see [decompose_OSLcurve](#) for details.

Value

The input table components [data.frame](#) will be returned with four additional columns: `$t.start`, `$t.end` defining the time intervals and `$ch.start`, `$ch.end` assigning those intervals to channel indices. If a [numeric](#) vector is given as input, a new [data.frame](#) will be returned.

Last updates

2020-08-23, DM: Replaced previous maximum searching algorithm with [DEoptim::DEoptim](#) (**update may have changed analysis results**)

2020-10-29, DM: Added `parallel.computing` argument; enhanced roxygen documentation (*minor update*)

Author(s)

Dirk Mittelstraß, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Mittelstraß, D., 2019. Decomposition of weak optically stimulated luminescence signals and its application in retrospective dosimetry at quartz (Master thesis). TU Dresden, Dresden.

Storn, R., Price, K., 1997. Differential Evolution – A Simple and Efficient Heuristic for global Optimization over Continuous Spaces. *Journal of Global Optimization* 11, 341–359.

See Also

[decompose_OSLcurve](#), [RLum.OSL_decomposition](#), [DEoptim::DEoptim](#), [fit_OSLcurve](#)

Examples

```
A <- optimise_OSLintervals(c(2, 0.5, 0.02), channel.width = 0.1, channel.number = 200)
print(A, row.names = FALSE)
```

plot_OSLcurve

Advanced plot function for component resolved CW-OSL curves

Description

This function is used for plotting CW-OSL curves and its signal components. It can handle data returned by [fit_OSLcurve](#) or [decompose_OSLcurve](#). Besides CW-OSL curves, pseudoLM-OSL curves and residual plots can also be plotted.

Usage

```
plot_OSLcurve(
  curve = NULL,
  components = NULL,
  display = "detailed",
  show.legend = TRUE,
  show.intervals = FALSE,
  show.crossec = FALSE,
  show.initial = FALSE,
  theme.set = ggplot2::theme_classic(),
  title = NULL,
  hide.plot = FALSE,
  filename = NULL
)
```

Arguments

curve	data.frame or matrix or RLum.Data.Curve (<i>optional</i>): CW-OSL curve x-Axis: \$time or first column as measurement time (must have constant time intervals); y-Axis: \$signal or second column as luminescence signal. Other columns will be plotted as component curves, in case no input object components is defined. If no input is given, a CW-OSL curve will be simulated with the parameters of components
components	data.frame (<i>optional</i>): Table with OSL component parameters. The parameters are used to approximate separate signal decay curves for each component. Need to have at least the columns: \$names, \$lambda and \$n. If an insufficient or no input object is provided, the ‘curve‘ object will be searched for component-related signal values.

display	character (<i>with default</i>): Sets the arrangement of graphs, see section Details .
show.legend	logical (<i>with default</i>): Draws a legend in the top right corner of the first graph.
show.intervals	logical (<i>with default</i>): Draws vertical lines into the residual plot showing the signal bin intervals (if available) for the CW-OSL decomposition with decompose_OSLcurve .
show.crossec	logical (<i>with default</i>): Displays photoionisation cross section values in the component table (if available).
show.initial	logical (<i>with default</i>): Displays signal share at the first channel in the component table (if available).
theme.set	ggplot2 object (<i>with default</i>): Graphical theme of the output plot. This argument is forwarded to <code>ggplot2::theme_set</code> . Recommended themes are <code>ggplot2::theme_minimal()</code> , <code>ggplot2::theme_classic()</code> and <code>ggplot2::theme_bw()</code> , see ggplot2::theme_bw or here for a full list.
title	character (<i>with default</i>): Plot title. Overwrites automatic titles but affects just the first (upper left) graph in case of multi-graph display setting. Set <code>title = NULL</code> for auto-title and <code>title = ""</code> for no title.
hide.plot	logical (<i>with default</i>): If true, plot is not drawn but can still be saved as file or caught by <code>A <- plot_OSLcurve(...)</code> . If caught, the plot can be drawn manually for example by using gridExtra::grid.arrange .
filename	character (<i>optional</i>): File name or path to save the plot as image. If just a file name is given, the image is saved in the working directory. The image type is chosen by the file ending. Both, vector images as well as pixel images are possible. Allowed are <code>.pdf</code> , <code>.eps</code> , <code>.svg</code> (vector graphics), <code>.jpg</code> , <code>.png</code> , <code>.bmp</code> (pixel graphics) and more, see ggplot2::ggsave .

Details

Change graph types with parameter: display

"detailed"	(default) Output plot consists of: Linear CW-OSL plot, pseudoLM-OSL plot, residual curve and component table
"lin"	Linear CW-OSL plot only
"compare_lin"	Linear CW-OSL plot with residual curve below and component table on bottom. Useful if two CW-OSL measurements are compared
"log"	CW-OSL plot with logarithmic y-Axis and linear x-Axis
"compare_log"	CW-OSL plot with logarithmic y-Axis with residual curve below and component table on bottom. Useful if two CW-OSL measurements are compared
"loglog"	Double-logarithmic CW-OSL plot
"LM"	PseudoLM-OSL plot
"res"	Plot of residual curve: Measurement minus fitting model
"tab"	Table of component parameters as image
"raw"	Raw x-y plot without further data

PseudoLM-OSL curves are created using the transformation described by Bulur (2000). The stimulation ramp duration is twice the CW-OSL duration. See Bos and Wallinga (2012) for a detailed explanation and discussion.

Value

An invisible `ggplot2::ggplot` object containing the diagram will returned. "Invisible" means, the no value will be returned (e.g. no console printout) if the function is not assigned to a variable via `<-`. If the function is assigned, the returned object can be further manipulated by `ggplot2`-package methods or manually drawn by various functions like for example `gridExtra::grid.arrange`.

Last update

2021-03-29, DM: Hidden output objects are now `ggplot2` objects if the plot is not a composite diagram

Author(s)

Dirk Mittelstraß, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Bos, A. J. J. and Wallinga, J., 2012. How to visualize quartz OSL signal components, *Radiation Measurements*, 47(9)

Bulur, E., 2000. A simple transformation for converting CW-OSL curves to LM-OSL curves, *Radiation Measurements*, 32(2)

See Also

[fit_OSLcurve](#), [\[RLum.OSL_decomposition\]](#), [RLum.OSL_global_fitting](#), [simulate_OSLcomponents](#)

Examples

```
# Set some arbitrary decay parameter for a dim CW-OSL measurement of quartz
components <- data.frame(name = c("fast", "medium", "slow"),
                        lambda = c(2, 0.5, 0.02),
                        n = c(1000, 1000, 10000))

# Simulate a CW-OSL curve including some signal noise
curve <- simulate_OSLcomponents(components, simulate.curve = TRUE, add.poisson.noise = TRUE)

# Display the simulated curve
plot_OSLcurve(curve, components)
```

plot_PhotoCrosssections

Plot comparison of CW-OSL component photoionisation cross sections of different models

Description

This function takes the `output.complex = TRUE` output of [fit_OS�curve](#) and draws the photoionisation cross sections of different models in relation to each other. If a stimulation wavelength between 465 and 480 nm was chosen, the photoionisation cross sections are also set in relation to literature values from Singarayer and Bailey (2003), Jain et al. (2003) and Durcan and Duller (2011).

Usage

```
plot_PhotoCrosssections(
  fit.list,
  stimulation.intensity = NULL,
  stimulation.wavelength = NULL,
  K.selected = NULL,
  title = NULL,
  hide.plot = FALSE,
  filename = NULL
)
```

Arguments

<code>fit.list</code>	list (required) : Output object of fit_OS�curve . The object must be created with the setting <code>output.complex = TRUE</code> .
<code>stimulation.intensity</code>	numeric (optional) : Intensity of optical stimulation in mW/cm^2 . Used to calculate the photoionisation cross sections. If not given, the input value for fit_OS�curve is used
<code>stimulation.wavelength</code>	numeric (optional) : Wavelength of optical stimulation in nm. Used to calculate the photoionisation cross sections. If not given, the input value for fit_OS�curve is used
<code>K.selected</code>	numeric (optional) : Draws a red rectangle around the <code>K = K.selected</code> row, thus highlighting the model of choice.
<code>title</code>	character (with default) : Plot title. Set <code>title = NULL</code> for no title.
<code>hide.plot</code>	logical (with default) : If true, plot is not drawn but can still be saved as file or caught by <code>A <- plot_PhotoCrosssections(...)</code> . If caught, the plot can be drawn manually for example by using gridExtra::grid.arrange .
<code>filename</code>	character (optional) : File name or path to save the plot as image. If just a file name is given, the image is saved in the working directory. The image type is chosen by the file ending. Both, vector images as well as pixel images are possible. Allowed are <code>.pdf</code> , <code>.eps</code> , <code>.svg</code> (vector graphics), <code>.jpg</code> , <code>.png</code> , <code>.bmp</code> (pixel graphics) and more, see ggplot2::ggsave .

Details

The photoionisation cross section ranges of the reference components are defined as following:

Component	Lower limit (cm ²)	Upper limit (cm ²)
Ultrafast	1e-16	1e-15
Fast	1.9e-17	3.1e-17
Medium	3e-18	9e-18
Slow1	1e-18	1.85e-18
Slow2	1.1e-19	4e-19
Slow3	1e-20	4.67e-20
Slow4	1e-21	1e-20

Value

An invisible `ggplot2::ggplot` object containing the diagram will returned. "Invisible" means, the no value will be returned (e.g. no console printout) if the function is not assigned to a variable via `<-`. If the function is assigned, the returned object can be further manipulated by `ggplot2-package` methods or manually drawn by various functions like for example `gridExtra::grid.arrange`.

Last updates

2020-11-04, DM: Added roxygen documentation

Author(s)

Dirk Mittelstraß, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Durcan, J.A., Duller, G.A.T., 2011. The fast ratio: A rapid measure for testing the dominance of the fast component in the initial OSL signal from quartz. *Radiation Measurements* 46, 1065–1072.

Jain, M., Murray, A.S., Bøtter-Jensen, L., 2003. Characterisation of blue-light stimulated luminescence components in different quartz samples: implications for dose measurement. *Radiation Measurements* 37, 441–449.

Singarayer, J.S., Bailey, R.M., 2003. Further investigations of the quartz optically stimulated luminescence components using linear modulation. *Radiation Measurements, Proceedings of the 10th international Conference on Luminescence and Electron-Spin Resonance Dating (LED 2002)* 37, 451–458.

See Also

[fit_OSLcurve](#), [RLum.OSL_global_fitting](#)

Examples

```
# Set some arbitrary decay parameter for a dim CW-OSL measurement of quartz
name <- c("fast", "slow")
lambda <- c(2, 0.02)
n <- c(1e6, 5e7)

# Build a component table
components <- data.frame(name, lambda, n)

# Simulate the CW-OSL curve and add some signal noise
curve <- simulate_OSLcomponents(components, simulate.curve = TRUE, add.poisson.noise = TRUE)

# Perform nonlinear regression at the simulated curve
fit_results <- fit_OSLcurve(curve, K.max = 2, output.complex = TRUE)

# Plot the fitting iterations and set them into context
plot_PhotoCrosssections(fit_results)
```

RLum.OSL_correction *Check and correct CW-OSL curves in RLum.Analysis data sets*

Description

CW-OSL measurements are often affected by background signals or might be measured under inconsistent detection settings. This function provides tools to test and solve some common problems.

Usage

```
RLum.OSL_correction(
  object,
  record_type = "OSL",
  background_sequence = NULL,
  subtract_offset = 0,
  check_consistency = TRUE,
  remove_light_off = TRUE,
  limit_duration = 20,
  PMT_pulse_pair_resolution = 18,
  verbose = TRUE
)
```

Arguments

object [RLum.Analysis](#) or list of [RLum.Analysis](#) (**required**): Data set of one or multiple CW-OSL measured aliquots.

record_type	character (<i>with default</i>): Type of records selected from the input object, see <code>object[[[]]]@records[[[]]]@recordType</code> . Common are: "OSL", "SGOSL" or "IRSL".
background_sequence	numeric vector (<i>optional</i>): Indices of list items with CW-OSL measurements of empty aliquots. The records in these list items are used to calculate one average CW-OSL background curve with <code>sum_OSLcurves</code> . This background curve is subtracted from each CW-OSL record of the data set. The attributes <code>@recordType</code> of the background measurements will be renamed to "{record_type}background".
subtract_offset	numeric (<i>optional</i>): Signal offset value in counts per second (<i>cts/s</i>). Value is handled as background level and will be subtracted from each CW-OSL record.
check_consistency	logical (<i>with default</i>): The CW-OSL component identification and separation procedure requires uniform detection parameters throughout the whole data set. If TRUE, all records are compared for their channel width and their number of channels. Those records with the most frequent set of channel parameters keep their <code>@recordType</code> attribute, while records with other sets of measurement parameter will be enumerated <code>record_type "{record_type}2", "{record_type}3"</code> and so on.
remove_light_off	logical (<i>with default</i>): Checks if the records contain zero-signal intervals at beginning and/or end of the measurement and removes them. Useful to tailor single-grain measurements.
limit_duration	numeric (<i>with default</i>): Reduce measurement duration to input value in seconds (<i>s</i>). Long measurement duration can lead to over-fitting at the component identification of Step 1 which may induce systematic errors, see Mittelstrass (2019). Thus, limiting the OSL record length ensures sufficient accuracy regarding the Fast and Medium component analysis. If however, slow decaying components are of interest, <code>limit_duration = NA</code> is recommended.
PMT_pulse_pair_resolution	numeric (<i>with default</i>): Time span of the pulse-pair resolution of the PMT in nanoseconds (<i>ns</i>). If a value is given, the signal values will be corrected for time-resolution related non-linearity at height counting rates, see <i>Details</i> . Set <code>PMT_pulse_pair_resolution = NA</code> if algorithm shall be omitted.
verbose	logical (<i>with default</i>): Enables console text output.

Details

This function processes data sets created within the [Luminescence-package](#) (Kreutzer et al. 2012). Those data sets must be formatted as `RLum.Analysis` objects. Output objects will also be `RLum.Analysis` objects and are meant for further analysis with `RLum.OSL_global_fitting`.

The data preparation tools are executed in the following order:

1. `check_consistency`
2. `remove_light_off`
3. `limit_duration`
4. `PMT_pulse_pair_resolution`

5. background_sequence
6. subtract_offset

Currently, not all functions are available.

Details to remove_light_off: The algorithm does the following: (1) Create global reference curve with [sum_OSLcurves](#) (2) Search for the maximum in the first half of the reference curve and remove all data points before the maximum . Do this for all curves of the selected 'record_type'. (3) Search for an inflection point with negative curvature (minimum of second differential) in the second half of the reference curve. If the next data point has at least 50% less signal, remove all data points after the inflection point. Do this for all curves of the selected 'record_type'.

Details to PMT_pulse_pair_resolution:

The algorithm corrects non-linearity of signal values due to insufficient pulse-pair resolution of the photo-multiplier tube (PMT). Equation (6-2) of the *Hamamatsu Photomultiplier Handbook* is used:

$$I_{corrected} = I_{measured} / (1 - I_{measured} * resolution)$$

The algorithm does not account for PMT saturation and PMT aging effects. As default pulse-pair resolution 18 ns is pre-defined, the *Hamamatsu H7360* series pulse-pair resolution according to the data sheet. The H7360-02 is the default PMT in *Freiberg Instruments lexsys OSL/TL* readers. *DTU Physics Risoe TL/OSL* reader deploy *ET Enterprise 9235B* series PMTs as default. For these PMTs, the pulse-pair resolutions is not given in the data sheets and relies on the operation voltage. However, due to the pulse properties given in the data sheets, it is unlikely that those PMTs have a better pulse-pair resolution than 18 ns.

Impact of a pulse-pair resolution correction of 18 ns

Measured signal	Corrected signal	Signal underestimation
1000 cts/s	1000 cts/s	0.00 %
10000 cts/s	10002 cts/s	0.02 %
50000 cts/s	50045 cts/s	0.09 %
100000 cts/s	100180 cts/s	0.18 %
500000 cts/s	504541 cts/s	0.91 %
1000000 cts/s	1018330 cts/s	1.83 %

Value

The input object, a [list](#) of [RLum.Analysis](#) objects, is given back with eventual changes in the elements `object[[]]@records[[]]@recordType` and `object[[]]@records[[]]@data`.

The returned data set contains a new list element `object[["CORRECTION"]]` which provides a [list](#) of the input parameters and additional data depending on the applied tools.

Last updates

2022-01-02, DM: Revised PMT_pulse_pair_resolution algorithm.

Author(s)

Dirk Mittelstrass, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Hamamatsu, 2007. Photomultiplier Tubes: Basics and Applications, Third Edition (Edition 3A). Hamamatsu Photonics K. K., Hamamatsu City.

Kreutzer, S., Schmidt, C., Fuchs, M.C., Dietze, M., Fischer, M., Fuchs, M., 2012. Introducing an R package for luminescence dating analysis. *Ancient TL*, 30 (1), 1-8.

Mittelstraß, D., 2019. Decomposition of weak optically stimulated luminescence signals and its application in retrospective dosimetry at quartz (Master thesis). TU Dresden, Dresden.

See Also

[RLum.OSL_global_fitting](#), [RLum.OSL_decomposition](#), [sum_OSLcurves](#)

Examples

```
# 'FB_10Gy' is a dose recovery test with the Fontainebleau quartz
# measured with a lexyg research with green LED stimulation
data_path <- system.file("examples", "FB_10Gy_SAR.bin", package = "OSLdecomposition")
data_set <- Luminescence::read_BIN2R(data_path, fastForward = TRUE)

# To correct for the background signal, subtracted the average curve from the
# OSL curves of an empty aliquot (list item 11) from all other OSL records:
data_set_corrected <- RLum.OSL_correction(data_set, background = 11, remove_light_off = FALSE)

# Plot background corrected global average CW-OSL curve
sum_OSLcurves(data_set_corrected, output.plot = TRUE, record_type = "OSL")

# Plot background curve
sum_OSLcurves(data_set_corrected, output.plot = TRUE, record_type = "OSLbackground")
```

Description

Calculates the CW-OSL signal component intensities for each CW-OSL measurement under the requirement that the decay rates are already given. The signal decomposition process uses an analytical approach described in detail in Mittelstrass (2019) and Mittelstrass et al. (in preparation). This function processes [RLum.Analysis](#) data sets created within the [Luminescence-package](#) (Kreutzer et al. 2012).

Usage

```
RLum.OSL_decomposition(
  object,
  record_type = "OSL",
  K = 3,
  decay_rates = NULL,
  report = FALSE,
  report_dir = NULL,
  image_format = "pdf",
  open_report = TRUE,
  rmd_path = NULL,
  verbose = TRUE
)
```

Arguments

object	RLum.Analysis or list of RLum.Analysis (required): Data set of one or multiple CW-OSL measured aliquots. The data set must either contain a list element <code>\$OSL_COMPONENTS</code> or the parameter <code>decay_rates</code> must be defined.
record_type	character (<i>with default</i>): Type of records, selected by the RLum.Analysis attribute <code>@recordType</code> . Common are: "OSL", "SGOSL" or "IRSL".
K	numeric (<i>with default</i>): Number of components. Selects the according result table in the <code>\$OSL_COMPONENTS</code> list item of the data set object.
decay_rates	numeric vector or data.frame (<i>optional</i>): User-defined component decay rates. If this parameter is defined, the parameter K will be ignored. If the input object is a data.frame , then the decay rates must be stored in the column <code>\$lambda</code> .
report	logical (<i>with default</i>): Creates a html report, saves it in the <code>report_dir</code> directory. The report contains the results and detailed information on the data processing.
report_dir	character (<i>optional</i>): Path of output directory if <code>report = TRUE</code> . If <code>report_dir = NULL</code> (default), a temporary folder is used which is deleted when the R session is closed. File paths are also allowed as parameter, then a new directory named after the OSL data file will be created.
image_format	character (<i>with default</i>): Image format of the automatically saved graphs if <code>report = TRUE</code> and <code>report_dir</code> is set. Allowed are <code>.pdf</code> , <code>.eps</code> , <code>.svg</code> (vector graphics), <code>.jpg</code> , <code>.png</code> , <code>.bmp</code> (pixel graphics) and more, see ggplot2::ggsave . The images are saved in the <code>report_dir</code> subfolder <code>/report_figures</code> . Set <code>image_format = NULL</code> if no images shall be saved.

open_report	logical (<i>with default</i>): If set to TRUE a browser window displaying the report will be opened automatically.
rmd_path	character (<i>with default</i>): For advanced users: File path to the <code>rmarkdown</code> source code file of the report. This allows to execute a manipulated version of the report.
verbose	logical (<i>with default</i>): Enables console text output.

Details

The workflow of this function is as follows:

1. `optimise_OSLintervals`: Approximates the optimal integration intervals. Uses the global average curve as time axis template. If none global average curve is given, one is automatically created using `sum_OSLcurves`.
2. `decompose_OSLcurve`: Calculates component intensities for **all** `record_type` measurements. Uses the "det" algorithm if a background correction was performed with `RLum.OSL_correction` or the "det+nls" algorithm if no background correction was performed. For error estimation, the "empiric" approach is used.
3. Creates a html report to summarize the results (*optional*).

Data sets must be formatted as `RLum.Analysis` objects and should have been processed with `RLum.OSL_correction` and `RLum.OSL_global_fitting` beforehand. Output objects are also `RLum.Analysis` objects and are meant for equivalent dose determination with `Luminescence::analyse_SAR.CWOSL`.

If `report = TRUE`, a html report of the results is rendered by the `rmarkdown-package` and saved in the working directory, which is usually the directory of the data file. This report can be displayed, shared and published online without any requirements regarding the operation system or installed software. However, an internet connection is needed to display the *MathJax* encoded equations and special characters. The *Rmarkdown* source code of the report can be found with the following command:

```
system.file("rmd", "report_Step2.Rmd", package = "OSLdecomposition")
```

Value

The input object, a list of `RLum.Analysis` objects is returned but with a new list element object `[["DECOMPOSITION"]]`, containing:

- `$decomposition.input data.frame`: Set of input components. Relevant is just the column `$lambda`
- `$results data.frame`: Overview table of decomposition
- `$parameters list`: Input and algorithm parameters

The `RLum.Data.Curve` attribute `@info` of each CW-OSL record contains the new entry `$COMPONENTS` with the curve-individual signal component parameters. It can be read for example by:

```
object[[i]]@records[[j]]@info[["COMPONENTS"]]
```

Last updates

2022-05-02, DM: Added new parameter `open_report` to give control over automatic browser opening

Author(s)

Dirk Mittelstrass, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Kreutzer, S., Schmidt, C., Fuchs, M.C., Dietze, M., Fischer, M., Fuchs, M., 2012. Introducing an R package for luminescence dating analysis. *Ancient TL*, 30 (1), 1-8.

Mittelstraß, D., 2019. Decomposition of weak optically stimulated luminescence signals and its application in retrospective dosimetry at quartz (Master thesis). TU Dresden, Dresden.

See Also

[RLum.OSL_global_fitting](#), [decompose_OSLcurve](#), [optimise_OSLintervals](#), [Luminescence::analyse_SAR.CWOSL](#)

Examples

```
#'FB_10Gy' is a dose recovery test with the Fontainebleau quartz
# measured in a lexsyg research with green LED stimulation
data_path <- system.file("examples", "FB_10Gy_SAR.bin", package = "OSLdecomposition")
data_set <- Luminescence::read_BIN2R(data_path, fastForward = TRUE)

# Separate components
data_set_decomposed <- RLum.OSL_decomposition(
  data_set, decay_rates = c(0.8, 0.05))
```

RLum.OSL_global_fitting

Identify CW-OSL signal components in RLum.Analysis data sets

Description

First, all CW-OSL records are combined to one global average CW-OSL curve, then the multi-exponential fitting approach of Bluszcz and Adamiec (2006) is applied. This function processes [RLum.Analysis](#) data sets created within the [Luminescence-package](#) (Kreutzer et al. 2012).

Usage

```

RLum.OSL_global_fitting(
  object,
  record_type = "OSL",
  K_maximum = 5,
  F_threshold = 150,
  stimulation_intensity = 35,
  stimulation_wavelength = 470,
  report = FALSE,
  report_dir = NULL,
  image_format = "pdf",
  open_report = TRUE,
  rmd_path = NULL,
  verbose = TRUE
)

```

Arguments

object	RLum.Analysis or list of RLum.Analysis (required): Data set of one or multiple CW-OSL measured aliquots.
record_type	character (<i>with default</i>): Type of records, selected by the RLum.Analysis attribute @recordType. Common are: "OSL", "SGOSL" or "IRSL".
K_maximum	numeric (<i>with default</i>): Maximum number of components K , see fit_OSLcurve .
F_threshold	numeric (<i>with default</i>): Fitting stop criterion, see fit_OSLcurve .
stimulation_intensity	numeric (<i>with default</i>): Intensity of optical stimulation in mW / cm^2 . Used to calculate photo-ionisation cross-sections, see fit_OSLcurve .
stimulation_wavelength	numeric (<i>with default</i>): Wavelength of optical stimulation in nm . Used to calculate photo-ionisation cross-sections, see fit_OSLcurve .
report	logical (<i>with default</i>): Creates a html report, saves it in the report_dir directory. The report contains the results and detailed information on the data processing.
report_dir	character (<i>optional</i>): Path of output directory if report = TRUE. If report_dir = NULL (default), a temporary folder is used which is deleted when the R session is closed. File paths are also allowed as parameter, then a new directory named after the OSL data file will be created.
image_format	character (<i>with default</i>): Image format of the automatically saved graphs if report = TRUE and report_dir is set. Allowed are .pdf, .eps, .svg (vector graphics), .jpg, .png, .bmp (pixel graphics) and more, see ggplot2::ggsave . The images are saved in the report_dir subfolder /report_figures. Set image_format = NULL if no images shall be saved.
open_report	logical (<i>with default</i>): If set to TRUE a browser window displaying the report will be opened automatically.

rmd_path	character (<i>with default</i>): For advanced users: File path to the rmarkdown source code file of the report. This allows to execute manipulated versions of the report.
verbose	logical (<i>with default</i>): Enables console text output.

Details

The workflow of this function is as follows:

1. [sum_OSLcurves](#): Combine all measurements of type `record_type` to one global average curve.
2. [fit_OSLcurve](#): Identify OSL components by a multi-exponential fitting.
3. Create a html report to summarize the results (*optional*).

Data sets must be formatted as [RLum.Analysis](#) objects and should have been processed with [RLum.OSL_correction](#) beforehand. Output objects are also [RLum.Analysis](#) objects and are meant for further analysis with [RLum.OSL_decomposition](#).

If `report = TRUE`, a html report of the results is rendered by the [rmarkdown-package](#) and saved in the working directory, which is usually the directory of the data file. This report can be displayed, shared and published online without any requirements to the operation system or installed software. However, an internet connection is needed to display the *MathJax* encoded equations and special characters. The *Rmarkdown* source code of the report can be found with the following command:

```
system.file("rmd", "report_Step1.Rmd", package = "OSLdecomposition")
```

Value

The input object, a `list` of [RLum.Analysis](#) objects is returned but with a new list element `object[["OSL_COMPONENTS"]]`, containing:

- `$decay.rates` **numeric** vector: Decay rates of F-test recommendation or last successful fitting.
- `$K.selected` **numeric**: Number of components of F-test recommendation or last successful fitting.
- `$F.test` **data.frame**: F-test table.
- `$F.test.print` **data.frame**: F-test table but formatted for console output and display with [knitr::kable](#).
- `$info.text` **list**: Short process log.
- `$component.tables` **list** of **data.frames**: Signal component tables for all curve models.
- `$curve` **list**: Global average curve created from all `record_type` curves in the data set.
- `$components` **data.frame**: Signal component table of F-test recommendation or last successful fitting.
- `$fit.results` **list**: Returned fitting objects of [DEoptim::DEoptim](#) and [minpack.lm::nlsLM](#) for all curve models.
- `$plot.data` **data.frame**: Model overview table for photo-ionisation cross-section plotting with [plot_PhotoCrosssections](#).
- `$parameters` **list**: Input and algorithm parameters.

Last updates

2022-05-02, DM: Added new parameter `open_report` to give control over automatic browser opening

Author(s)

Dirk Mittelstrass, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Bluszcz, A., Adamiec, G., 2006. Application of differential evolution to fitting OSL decay curves. *Radiation Measurements* 41, 886–891.

Kreutzer, S., Schmidt, C., Fuchs, M.C., Dietze, M., Fischer, M., Fuchs, M., 2012. Introducing an R package for luminescence dating analysis. *Ancient TL*, 30 (1), 1-8.

See Also

[RLum.OSL_correction](#), [RLum.OSL_decomposition](#), [sum_OSLcurves](#), [fit_OSLcurve](#)

Examples

```
# 'FB_10Gy' is a dose recovery test with the Fontainebleau quartz
# measured in a lexsyg research with green LED stimulation
data_path <- system.file("examples", "FB_10Gy_SAR.bin", package = "OSLdecomposition")
data_set <- Luminescence::read_BIN2R(data_path, fastForward = TRUE)

# Check data set and perform background correction
data_set_corrected <- RLum.OSL_correction(data_set,
  background = 11,
  remove_light_off = FALSE)

# Identify components
data_set_fitted <- RLum.OSL_global_fitting(
  data_set_corrected,
  K_maximum = 2,
  stimulation_intensity = 50,
  stimulation_wavelength = 530)
```

 simulate_OSLcomponents

Simulates signal component decay curves and whole CW-OSL curves

Description

This function builds a bulk CW-OSL curve and CW-OSL component decay curves from OSL component parameters. Therewith it supports [fit_OSLcurve](#), [decompose_OSLcurve](#) and [plot_OSLcurve](#) by providing model and residual curves.

Usage

```
simulate_OSLcomponents(
  components,
  curve = NULL,
  channel.width = 0.1,
  channel.number = 400,
  simulate.curve = FALSE,
  add.poisson.noise = TRUE,
  add.gaussian.noise = 0,
  add.background = 0,
  round.values = TRUE
)
```

Arguments

- | | |
|--------------------|--|
| components | data.frame (required) : Table with component parameters. The table requires columns \$names, \$lambda and \$n, see section Examples . |
| curve | data.frame (optional) : CW-OSL curve serving as template for the time axis. The input table requires a column \$time. If no input object is given or the object contains no column \$signal, simulate.curve will be set TRUE. |
| channel.width | numeric (optional) : Channel width in seconds. Necessary for curve simulation if curve is not given. |
| channel.number | numeric (optional) : Number of channels resp. data points. Necessary for curve simulation if curve is not given. |
| simulate.curve | logical (with default) : Decides if the bulk CW-OSL signal shall be calculated from the component parameter. If FALSE, the output curve will take over the column \$signal from the input curve. If TRUE, a new column \$signal will be created which is the sum of all component curves. |
| add.poisson.noise | logical (with default) : Adds poisson distributed shot noise to \$signal if simulate.curve = TRUE. |
| add.gaussian.noise | numeric (with default) : Standard deviation of the detector noise in <i>cts/s</i> , added to \$signal if simulate.curve = TRUE. |

add.background **numeric** (*with default*): signal background level in *cts/s*, added to `$signal` if `simulate.curve = TRUE`

round.values **logical** (*with default*): Rounds `$signal` values to integers if `simulate.curve = TRUE`.

Value

A **data.frame** of a CW-OSL curve with the columns: `$time`, `$signal`, `$residual`, `$sum` and a signal decay curve for each single component named after the entries in the column `components$names` of the input object.

Last updates

2020-10-30, DM: Renamed from *simulate_OSLcurve* to *simulate_OSLcomponents*; Renamed argument from *template.curve* to *curve*; Rewrote roxygen documentation

Author(s)

Dirk Mittelstraß, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

References

Mittelstraß, D., 2019. Decomposition of weak optically stimulated luminescence signals and its application in retrospective dosimetry at quartz (Master thesis). TU Dresden, Dresden.

See Also

[fit_OSLcurve](#), [decompose_OSLcurve](#), [plot_OSLcurve](#)

Examples

```
# Set some arbitrary decay parameter for a dim CW-OSL measurement of quartz
components <- data.frame(name = c("fast", "medium", "slow"),
                        lambda = c(2, 0.5, 0.02),
                        n = c(1000, 1000, 10000))

# Simulate the CW-OSL curve and add some signal noise
curve <- simulate_OSLcomponents(components, simulate.curve = TRUE, add.poisson.noise = TRUE)

# Display the simulated curve
plot_OSLcurve(curve, components)
```

sum_OSLcurves	<i>Combine RLum OSL records to one global average curve</i>
---------------	---

Description

This function adds up all CW-OSL records of the same type saved in [RLum.Analysis](#) objects and calculates the arithmetic mean signal from all records for each channel. This is useful to create global average curve with sufficient signal-to-noise ratio for OSL components identification with [fit_OSLcurve](#) or to create a signal background reference curve.

Usage

```
sum_OSLcurves(
  object,
  record_type = "OSL",
  aliquot_selection = NULL,
  offset_value = 0,
  verbose = TRUE,
  output.plot = FALSE,
  theme.set = ggplot2::theme_classic(),
  plot.first = FALSE,
  title = "default",
  filename = NULL
)
```

Arguments

object	RLum.Analysis or list of RLum.Analysis (required): Data set of one or multiple aliquots containing CW-OSL records.
record_type	character (<i>with default</i>): Type of records which are selected from the input object, for example: "OSL", "SGOSL" or "IRSL".
aliquot_selection	numeric vector (<i>optional</i>): Vector specifying the indices of elements (aliquots) of a list of RLum.Analysis objects which shall be included.
offset_value	numeric (<i>with default</i>): Signal offset (background) which will be subtracted from each record.
verbose	logical (<i>with default</i>): Enables console text output.
output.plot	logical (<i>with default</i>): returns a plot with all data points of all records and the average curve
theme.set	ggplot2 object (<i>with default</i>): sets the graphical theme of the output plot. See ggplot2::theme_bw for available themes
plot.first	logical (<i>with default</i>): Plot includes additional drawing of first record_type record of first object list element.
title	character (<i>with default</i>): Plot title. Set title = "default" for an automatically generated title. Set title = NULL for no title.

filename **character** (*optional*): File name or path to save the plot as image. If just a file name is given, the image is saved in the working directory. The image type is chosen by the file ending. Both, vector images as well as pixel images are possible. Allowed are .pdf, .eps, .svg (vector graphics), .jpg, .png, .bmp (pixel graphics) and more, see [ggplot2::ggsave](#).

Value

A **data.frame** of the average CW-OSL curve is returned, containing two columns: \$time and \$signal.

Last updates

2020-10-30, DM: Overworked plotting; Expanded roxygen documentation

Author(s)

Dirk Mittelstraß, <dirk.mittelstrass@luminescence.de>

Please cite the package the following way:

Mittelstraß, D., Schmidt, C., Beyer, J., Heitmann, J. and Straessner, A.: R package OSLdecomposition: Automated identification and separation of quartz CW-OSL signal components, *in preparation*.

See Also

[fit_OSLcurve](#), [RLum.OSL_correction](#), [RLum.OSL_global_fitting](#)

Examples

```
# 'FB_10Gy' is a dose recovery test with the Fontainebleau quartz
# measured in a lexsyg research with green LED stimulation
data_path <- system.file("examples", "FB_10Gy_SAR.bin", package = "OSLdecomposition")
data_set <- Luminescence::read_BIN2R(data_path, fastForward = TRUE)

# Give average CW-OSL curve back
average_curve <- sum_OSLcurves(data_set)
```

Index

* **package**
OSLdecomposition-package, 2

character, 4, 14, 16, 19, 22, 23, 25, 26, 30, 31

data.frame, 3, 4, 6, 8–13, 22, 23, 26, 28, 29, 31

decompose_OSLcurve, 3, 8, 10–14, 23, 24, 28, 29

DEoptim::DEoptim, 8, 10, 12, 13, 26
DEoptim::DEoptim.control, 8, 12

fit_OSLcurve, 3, 6, 7, 11, 13, 15–17, 25–31

ggplot2, 14, 15, 30
ggplot2-package, 15, 17
ggplot2::ggplot, 15, 17
ggplot2::ggsave, 14, 16, 22, 25, 31
ggplot2::theme_bw, 14, 30
ggplot2::theme_set, 14
gridExtra::grid.arrange, 14–17

knitr::kable, 26

list, 9, 10, 16, 18, 20, 22, 23, 25, 26, 30
logical, 4, 8, 12, 14, 16, 19, 22, 23, 25, 26, 28–30

Luminescence-package, 19, 22, 24
Luminescence::analyse_SAR.CWOSL, 23, 24

matrix, 3, 8, 11, 13
minpack.lm::nlsLM, 4–6, 8, 10, 26

nls, 10
numeric, 3–5, 8, 11, 12, 16, 19, 22, 25, 26, 28–30

optimise_OSLintervals, 3, 4, 6, 11, 23, 24
OSLdecomposition
(OSLdecomposition-package), 2
OSLdecomposition-package, 2

plot_OSLcurve, 10, 13, 28, 29
plot_PhotoCrosssections, 10, 16, 26

RLum.Analysis, 18–20, 22–26, 30
RLum.Data.Curve, 3, 8, 11, 13, 23
RLum.OSL_correction, 18, 23, 26, 27, 31
RLum.OSL_decomposition, 6, 10, 13, 21, 21, 26, 27
RLum.OSL_global_fitting, 15, 17, 19, 21, 23, 24, 24, 31

rmarkdown, 23, 26
rmarkdown-package, 23, 26

simulate_OSLcomponents, 15, 28
sum_OSLcurves, 3, 8, 10, 19–21, 23, 26, 27, 30

vector, 9